



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

### ATI Specialty Materials Lockport Operations

695 Ohio Street  
Lockport, NY 14094

Fulfills the requirements of

### ISO/IEC 17025:2017

In the field of

### TESTING

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 22 April 2023

Certificate Number: AT-1775



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**ATI Specialty Materials Lockport Operations**

695 Ohio Street  
Lockport, NY 14094

Erik Ricker 716-439-5336  
[erik.ricker@ATIMetals.com](mailto:erik.ricker@ATIMetals.com)

**TESTING**

Valid to: **April 22, 2023**


Certificate Number: **AT-1775**

**Chemical**

<b>Specific Tests and/or Properties Measured</b>	<b>Specification, Standard, Method, or Test Technique</b>	<b>Items, Materials or Product Tested</b>	<b>Key Equipment or Technology</b>
Nitrogen Analysis (0.0002 to 0.54) %	Combustion ASTM E1019	Ni, Co, Steel Alloys	Leco TC600, ON836
Carbon Analysis (0.001 to 0.8) %	Combustion ASTM E1019	Ni, Co, Steel Alloys	LECO CS600, CSLS600
Sulfur Analysis (0.0003 to 0.024) %	Combustion ASTM E1019	Ni, Co, Steel Alloys	LECO CS600, CSLS600
Oxygen Analysis (0.0002 to 0.04) %	Combustion ASTM E1019	Ni, Co, Steel Alloys	Leco TC600, ON836
Various Elements (0.001 to 99.5) %	XRF ASTM E322, E572, E1085, E2465	Ni, Co, Steel Alloys	Magix Fast, PW2606
Various Elements (5 to 1 000) mg/L or ppm	Spark-OES ASTM E415, E1086	Ni, Co, Steel Alloys	ARL4460, ARL3460

Note:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. AT-1775.



R. Douglas Leonard Jr., VP, PILR SBU

